



PRODUCT SPECIFICATION



Product: CEDERROTH EYE WASH STATION

Reference code: 721500



1. General description

An Eye Wash station containing two bottles of Cederroth Eye Wash and plastic and textile adhesive bandages. The products are protected against dust and moisture by transparent lids. Space for additional products is provided behind a lid with first aid instructions. The Eye Wash bottles are instantly opened by twisting in either direction. The Adhesive bandages (plasters) are placed in dispensers for easy access and theft-proof storage. A special key is provided for refill replacement.

2. Intended use

To be used in first aid situations, especially suitable for work places where there is a risk of burns.

3. Technical description

Product dimensions

(l x w x h) cm: 56 x 29 x 12

Material

First aid station	Green parts:	Polystyrene, ABS-plastic
	Transparent lid:	Polycarbonate
	Sealing profiles:	EPDM rubber (synthetic rubber)
First aid products		Refer to Product specifications for each product

4. Composition

Consumer package

<u>Ref#</u>	<u>Product/Material</u>	<u>Pcs</u>
51616970	Eye Wash station (empty)	1
725200	Cederroth Eye Wash, 500 ml 2-p	1
6036	Salvequick Plastic Plasters in dispenser	1
6444	Salvequick Textile Plasters in dispenser	1
51616418	Eye Wash instruction (label)	1
51616620	Key to the dispensers	1
51616668	Refill instruction with drill template	1

Transport package

Number of consumer packages in transport package:	1
Dimensions (l x w x h) cm:	57 x 29 x 12
Weight (kg):	3,5



PRODUCT SPECIFICATION



5. Labeling

On product

Lot number and serial number
Instructions
Reference to the products

On consumer/transport package

Cederroth address, telephone- and
faxnumber, web-address
Product name, Reference code
Lot-number, Expiry date, EAN-code
Quantity.

6. Storage conditions

The products in the eye wash station should be stored in their original packages for maximum protection until time of use.

Shelf life: 5 years from production date for each component

7. Accessories

No accessories are necessary for the use of this product.

Authorization: QA&RA Mgr
2013.12.04 08:46:32
+01'00'